

Pensky Martens Flash Point Tester Closed Type

As per IP 34 ASTM D-93, IS 1448 (P-21)

For determination of flash point of petroleum products and mixture having a flash point below 49°C (120° F.)

The apparatus consists of cast iron stove fitted with a polished brass cover, a brass cup with cover which contains the test flame arrangement, stirrer (hand operated) and heating arrangement.

For electricity heated model a rectangular heater with an energy regulator fitted in a box.

Ref.: 515

